

Correspondence

Psychological Medicine, 46 (2016).

doi:10.1017/S0033291716001926

First published online 30 August 2016

Brief history of transcranial direct current stimulation (tDCS): from electric fishes to microcontrollers

Electrical stimulation to treat medical conditions is not a new therapy; it has been used to treat diseases for centuries. The first electricity sources used for electrical stimulation were produced by animal electricity. Antique Egyptians knew about the electrical properties of Nile catfish, but it is unclear if (and how) they experimented with them for clinical purposes. The first reported evidence of electrical stimulation arrives some centuries later in antique Greece times, when Plato and Aristotle described the ability of the torpedo fish to generate curative effects by its electric discharges (Althaus, 1873; Rockwell, 1896; Harris, 1908).

The first evidence of transcranial stimulation in history comes in Roman Empire times, when Scribonius Largus (the physician of the Roman Emperor Tiberius Claudius Nero Caesar) described how placing a live torpedo fish over the scalp could relieve headache in a patient (Scribonius Largus, 1529). Perhaps the first person known to have been cured by torpedo fish electricity was Anthero, a freed slave of Tiberius Caesar, who suffered *gutta* (probably gout) (Cambridge, 1977). In the late 11th century, the Muslim physician in Persia, Ibn-Sidah also suggested the use of torpedo fishes to treat epilepsy (Priori, 2003), placing the live catfish on the brows of the patients (Delbourgo, 2006). The use of electric fish stimulation also spread to Africa, where Jesuit missionaries in early modern Abyssinia (Ethiopia) reported that locals used catfishes as a method of expelling ‘Devils out of the human body’ (Delbourgo, 2006). Fish electricity was maybe the most popular type of electrical stimulation for more than 10 centuries though it is not clear how the effects were measured.

In 1660 the German scientist Otto von Guericke invented the first electrostatic generator (Comroe & Dripps, 1976), a frictional crank-controlled machine. This device could be considered the first stimulator device and its variations were used later by scientists like the Italian anatomist Leopoldo Marco Antonio Caldani in 1756 to stimulate muscles in sheep and frogs (Caldani, 1760). The Middlesex Hospital (England) was probably the first hospital to purchase an electrostatic machine in 1767 (Cambridge, 1977).

In 1745 Ewald Georg von Kleist invented the Leyden jar, the first capacitor in history (Keithley, 1999). This device could store electric charge produced from an electrostatic generator. Experimenters, like Anton de Haen in 1755 (Priestley, 1767) and Benjamin Franklin in 1757 (Franklin, 1757), were able to combine electrostatic generators and the Leyden jar for therapeutic electrification (McWhirter *et al.* 2015).

In 1773 the anatomist and physiologist John Hunter studied the torpedo fish thoroughly. These investigations were undertaken at the request of John Walsh, who showed that the ‘shocks’ produced by the torpedo fish were caused by the generation of electricity (Walsh, 1773). These kind of animals or fishes have an electric organ that, on brain command, generate a three-dimensional dipole field around their bodies, discharging single-cycle pulses from below 1 Hz to about 65 Hz at rest (Hopkins, 2009). Electric fish electricity is not direct current (DC); nevertheless it is the first reported kind of stimulation in history.

Unlike fish electricity and electrostatic electricity, DC is the flow of electrical charge that does not vary with time, generating a constant signal (Belove & Drossman, 1976). The birth of the DC generator was in the 1st century BC with the so-called Baghdad battery, attributed to the ancient Persian civilization (Scrosati, 2011), but other references attribute the invention to the Parthians, calling it the Parthian galvanic cell (Keyser, 1993). This invention remained forgotten until the 20th century, when the archeologist Wilhelm Köning discovered it in Iraq and it was possibly used for medical purposes. In the 18th century Luigi Galvani invented a DC battery (galvanic battery) and his nephew, Giovanni Aldini, was one of the first persons to utilize DC for clinical applications. Aldini’s most detailed account of DC clinical treatment concerns Luigi Lanzarini, a 27-year-old farmer suffering from melancholy madness (major depression), who had been committed to Santo Orsola Hospital, in Bologna, Italy on 17 May 1801, but first assessing the effects of galvanic currents on his own head (Fitzgerald, 2014). The patient’s mood progressively improved so that Lazarini was apparently completely cured several weeks after the beginning of the treatment (Parent, 2004).

Aldini’s work was the milestone which began the era of DC stimulation for neurological and psychiatric conditions. Later in 1802, Hellwag and Jacobi reported the use of transcranial DC, also reporting the first evidence of phosphenes using DC (Hellwag & Jacobi, 1802; Paulus, 2010). Around 1880 the application of brain

stimulation treatments on patients was particularly popular among German psychiatrists, pioneers in electrotherapy, an early tDCS method. Wilhelm Tiggas, Rudolph Gottfried Arndt (Steinberg, 2013a) and Erb (1883) tried to establish clear rules on the most beneficial application methods and doses in order to investigate which results it may produce and under what circumstances (Steinberg, 2014). The experimental designs with larger groups in electric therapy research protocols were a common factor in this age. For example, Arndt used 12 psychotic patients in his 1870 experiment. Despite being very detailed, Arndt's reports do not provide exact data about the strength of the applied current. Due to controversial reports (some with positive results and others with negative) and the lack of understanding of operating principles, electrotherapy was repeatedly suspected of attaining result through suggestion only (Steinberg, 2013a). Many other researchers used DC for the treatment of mental disorders during the 19th century and the early part of the 20th century, but the variation of procedures, unclear descriptions, few qualitative details and the misunderstood effect of polarization led to variable and/or inconclusive results. The use of DC stimulation was abandoned from the 1930s (Steinberg, 2013b).

In 1957 DC reappeared in electrosleep therapy and around 1960–1963 electro-anesthesia research incorporated DC bias. In 1964, motivated by animal studies that reported lasting changes in excitability using prolonged scalp DC stimulation, Lippold and Redfearn used 50–500 μA DC currents in 32 healthy subjects, and reported that anodal current induced an increase in alertness, mood and motor activity, whereas cathodal polarization induced quietness and apathy (Lippold & Redfearn, 1964; Guleyupoglu *et al.* 2013). Despite several follow-up research works, from the 1970s DC stimulation was once again abandoned, probably due to the introduction of new psychiatric drugs (Dubljević *et al.* 2014).

It was not until 1998 when the usage of DC was advocated and the modern tDCS was born, when Priori and his colleagues investigated the influence of DC in the brain by testing its effects on cerebral cortex excitability using transcranial magnetic stimulation (Brunoni *et al.* 2012). Characteristics of tDCS, such as the fact that it is non-invasive, mostly well tolerated and its mild adverse effects, have sparked great interest and increase in clinical studies recently (Brunoni *et al.* 2012).

The transcranial DC stimulators have evolved from a simple galvanic battery in the 18th and 19th centuries, passing from vacuum tubes and transistors to microprocessors and microcontroller technologies in the 20th century. Progress in microcontroller technology has enabled electronic and biomedical engineers to

build precise tDCS devices with a better control of stimulation parameters (Paulus & Opitz, 2013) at reduced costs. The future tDCS designs will focus on obtaining simple systems by way of reducing size, power consumption, weight and enhanced portability (Kouzani *et al.* 2016).

The tDCS is a promising tool for basic neuroscientists, clinical neurologists and psychiatrists in their quest to causally probe cortical representations of sensorimotor and cognitive functions, to facilitate the treatment of various neuropsychiatric disorders (Schlaug & Renga, 2008) and enhance neurological functions in healthy humans (Dubljević *et al.* 2014). Although tDCS has demonstrated benefits, the scientific community must communicate carefully about their findings, by providing neutral data to the media and public, since the popular media sometimes consider tDCS as a 'miracle device' (Riggall *et al.* 2015). Such sensationalistic news about the benefits of tDCS leads people to self-administer stimulation, as we can see in some Internet do-it-yourself tDCS forums (Wexler, 2015).

Acknowledgements

The authors wish to thank Jorge Barraza-Díaz and Gerardo Reyes-Acevedo for their technical support.

Declaration of Interest

None.

References

- Althaus J (1873). *A Treatise on Medical Electricity*. Longman: London.
- Belove C, Drossman MM (1976). *Systems and Circuits for Electrical Engineering Technology*. McGraw-Hill: New York.
- Brunoni AR, Nitsche MA, Bolognini N, Bikson M, Wagner T, Merabet L, Edwards DJ, Valero-Cabre A, Rotenberg A, Pascual-Leone A, Ferrucci R, Priori A, Boggio PS, Fregni F (2012). Clinical research with transcranial direct current stimulation (tDCS): challenges and future directions. *Brain Stimulation* 5, 175–195.
- Caldani LM (1760). Letter to Haller, 1756. In *Mémoires sur les parties sensibles et irritables du corps animal [Memoirs of Sensitive and Irritable Parts of the Animal Body]*, vol. 3 (ed. A. Haller), pp. 143–144. M.M. Bousquet: Lausanne, Switzerland.
- Cambridge NA (1977). Electrical apparatus used in medicine before 1900. *Proceedings of the Royal Society of Medicine* 70, 635–641.
- Comroe Jr. JH, Dripps RD (1976). Scientific basis for the support of biomedical science. *Science* 192, 105–111.
- Delbourgo J (2006). *A Most Amazing Scene of Wonders: Electricity and Enlightenment in Early America*. Harvard University Press: Cambridge, MA.

- Dubljević V, Saigle V, Racine E** (2014). The rising tide of tDCS in the media and academic literature. *Neuron* **82**, 731–736.
- Erb W** (1883). *Handbook of Electrotherapeutics*. Wood: New York.
- Fitzgerald PB** (2014). Transcranial pulsed current stimulation: a new way forward? *Clinical Neurophysiology* **125**, 217–219.
- Franklin B** (1757). An account of the effects of electricity in paralytic cases. In a letter to John Pringle. *Philosophical Transactions of the Royal Society of London* **50**, 481–483.
- Guleyupoglu B, Schestatsky P, Edwards D, Fregni F, Bikson M** (2013). Classification of methods in transcranial electrical stimulation (tES) and evolving strategy from historical approaches to contemporary innovations. *Journal of Neuroscience Methods* **219**, 297–311.
- Harris W** (1908). *Electrical Treatment*. Cassell: London.
- Hellweg CF, Jacobi CWM** (1802). *Erfahrungen über die Heilkräfte des Galvanismus, und Betrachtungen über desselben chemische und physiologische Wirkungen [Experiences about the Healing Powers of Galvanism, and Considerations of the Same Chemical and Physiological Effects]*. Perthes: Hamburg, Germany.
- Hopkins CD** (2009). Electrical perception and communication. *Encyclopedia of Neuroscience* **3**, 813–831.
- Keithley JF** (1999). The Leyden jar – the first capacitor. In *The Story of Electrical and Magnetic Measurements: from 500 BC to the 1940s*, chapter 6. IEEE: Piscataway, NJ.
- Keyser PT** (1993). The purpose of the Parthian galvanic cells: a first-century AD electric battery used for analgesia. *Journal of Near Eastern Studies* **52**, 81–98.
- Kouzani AZ, Jaberzadeh S, Zoghi M, Usma C, Parastarfeizabadi M** (2016). Development and validation of a miniature programmable tDCS device. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **24**, 192–198.
- Lippold OCJ, Redfearn JWT** (1964). Mental changes resulting from the passage of small direct currents through the human brain. *British Journal of Psychiatry* **110**, 768–772.
- McWhirter L, Carson A, Stone J** (2015). The body electric: a long view of electrical therapy for functional neurological disorders. *Brain* **138**, 1113–1120.
- Parent A** (2004). Giovanni Aldini: from animal electricity to human brain stimulation. *Canadian Journal of Neurological Sciences* **31**, 576–584.
- Paulus W** (2010). On the difficulties of separating retinal from cortical origins of phosphenes when using transcranial alternating current stimulation (tACS). *Clinical Neurophysiology* **121**, 987–991.
- Paulus W, Opitz A** (2013). Ohm's law and tDCS over the centuries. *Clinical Neurophysiology* **124**, 429–430.
- Priestley J** (1767). *The History and Present State of Electricity with Original Experiments*. J. Dodsley, J. Johnson, B. Davenport, and T. Cadell: London, UK.
- Priori A** (2003). Brain polarization in humans: a reappraisal of an old tool for prolonged non-invasive modulation of brain excitability. *Clinical Neurophysiology* **114**, 589–595.
- Riggall K, Forlini C, Carter A, Hall W, Weier M, Partridge B, Meinzer M** (2015). Researchers' perspectives on scientific and ethical issues with transcranial direct current stimulation: an international survey. *Scientific Reports* **5**, article number 10618.
- Rockwell AD** (1896). *The Medical and Surgical Uses of Electricity*. Wood: New York.
- Schlaug G, Renga V** (2008). Transcranial direct current stimulation: a noninvasive tool to facilitate stroke recovery. *Expert Review of Medical Devices* **5**, 759–768.
- Scribonius Largus** (1529). *De Compositionibus Medicamentorum [The Compositions of Drugs]*. In *De Compositionibus Medicamentorum. Liber Unus* (ed. J. Ruello). Wechel: Paris, France. (8ll., 31 fol [ved. Fol. Ib, 3rd paragraph]).
- Scrosati B** (2011). History of lithium batteries. *Journal of Solid State Electrochemistry* **15**, 1623–1630.
- Steinberg H** (2013a). A pioneer work on electric brain stimulation in psychotic patients. Rudolph Gottfried Arndt and his 1870s studies. *Brain Stimulation* **6**, 477–481.
- Steinberg H** (2013b). Transcranial direct current stimulation (tDCS) has a history reaching back to the 19th century. *Psychological Medicine* **43**, 669–671.
- Steinberg H** (2014). 'Auch die Electricität leistet keine Wunder!' Die vergessenen Beiträge deutscher Psychiater um 1880 zur Therapie von Depressionen und Psychosen [Even electricity cannot work wonders! Neglected achievements by German psychiatrists around 1880 in the treatment of depressions and psychoses]. *Der Nervenarzt* **85**, 872–886.
- Walsh J** (1773). On the electric property of torpedo. In a letter to Benjamin Franklin. *Philosophical Transactions of the Royal Society of London* **63**, 461–481.
- Wexler A** (2015). The practices of do-it-yourself brain stimulation: implications for ethical considerations and regulatory proposals. *Journal of Medical Ethics* **42**, 211–215.

C. I. SARMIENTO^{1,2}, D. SAN-JUAN² AND V. B. S. PRASATH^{3,*}

¹Biomedical Engineering Laboratory, Division of Basic Sciences and Engineering, Metropolitan Autonomous University at Iztapalapa, Mexico City, Mexico

²Department of Clinical Research, National Institute of Neurology and Neurosurgery, Mexico City, Mexico

³Computational Imaging and VisAnalysis (CIVA) Laboratory, Department of Computer Science, University of Missouri, Columbia, MO, USA

* Author for correspondence: V. B. S. Prasath, Computational Imaging and VisAnalysis (CIVA) Laboratory, Department of Computer Science, University of Missouri, Columbia, MO 65211, USA.

(Email: surya.iit@gmail.com)